


Test Report

(Electronic version)

扫码下载报告

Verification Website: www.gtcc.net.cn

Verification Code: IPJP-8687-14



No:210188495

Issue Date:2021-07-06

Applicant: Shantou Green Beauty Industrial Co.,Ltd.

Address: Liangying Yongfeng Industrial Zone, Chaonan District, Shantou City

Information confirmed by applicant:

Disposable stick-on mask

Quantity: 30 pieces

Date of manufacture: 2021.6

Size: L0

Manufacture's name: Shantou Green Beauty Industrial Co.,Ltd.

Basis of judgement:

Q/STLM 001-2021 《一次性贴式口罩》

Date Received/Date Test Started: 2021-06-26

Conclusion:

Bacterial filtration efficiency	M
Coliform group	M
Total amount of bacterial colony	M
Total amount of fungus colony	M
Pseudomonas aeruginosa	M
Staphylococcus aureus	M
Hemolytic streptococcus	M
Airflow resistance	M

Note: "M"-Meet the standard's requirement "F"-Fail to meet the standard's requirement
"—"—No comment

Remark:

This report is the english translation version of the report 210188494.

The decision indicators are derived from the enterprise standard required by client (Q/STLM 001-2021).Our inspection capacity authorized by CMA covers the inspection items of Q/STLM 001-2021.

All the tested items are tested under the standard condition (except for indication).

Copies of the report are valid only re-stamped.

The experiment was carried out at No.1, Zhujiang Road, Panyu District, Guangzhou, Guangdong, P.R.China.

Approved By:

Yuan Liu Engineer

刘圆



Page 1 of 3

Test Report

(Electronic version)

No:210188495



Test Report

(Electronic version)

No:210188495

Bacterial filtration efficiency(%) (YY 0469-2011 Annex B, Test bacteria: Staphylococcus aureus ATCC 6538, Test area: 49cm ² , Flow rate: 28.3L/min, Mean particle size: 3.0 μm, Result of the positive control: 1.9×10 ³ CFU, Result of the negative control: <1CFU)	M
Requirement ≥95 (Q/STLM 001-2021)	
BFE ₁ 99.2	
BFE ₂ 99.0	
BFE ₃ 99.3	
Coliform group (GB 15979-2002 Annex B)	M
Requirement Must not be detected (Q/STLM 001-2021)	
Not detected	
Total amount of bacterial colony(CFU/g) (GB 15979-2002 Annex B)	M
Requirement ≤100 (Q/STLM 001-2021)	
<20	
Total amount of fungus colony(CFU/g) (GB 15979-2002 Annex B)	M
Requirement Must not be detected (Q/STLM 001-2021)	
Not detected	
Pseudomonas aeruginosa (GB 15979-2002 Annex B)	M
Requirement Must not be detected (Q/STLM 001-2021)	
Not detected	
Staphylococcus aureus (GB 15979-2002 Annex B)	M
Requirement Must not be detected (Q/STLM 001-2021)	
Not detected	
Hemolytic streptococcus (GB 15979-2002 Annex B)	M
Requirement Must not be detected (Q/STLM 001-2021)	
Not detected	
Airflow resistance(Pa/cm²) (YY/T 0969-2013 5.6)	M
Requirement ≤49 (Q/STLM 001-2021)	
1# 32.0	
2# 33.7	
3# 32.4	




———End of Report———

Page 3 of 3

Test Report

(Electronic version)

扫码下载报告



Verification Website: www.gttc.net.cn

Verification Code: FCYC-0336-44



No:210188493

Issue Date:2021-07-01

Applicant: Shantou Green Beauty Industrial Co.,Ltd.

Address: Liangying Yongfeng Industrial Zone, Chaonan District, Shantou City

Information confirmed by applicant:

Disposable stick-on mask

Quantity: 10 pieces

Date of manufacture: 2021.6

Size: L0

Manufacture's name: Shantou Green Beauty Industrial Co.,Ltd.

Basis of judgement:

Q/STLM 001-2021 《一次性贴式口罩》

Date Received/Date Test Started: 2021-06-26

Conclusion:

Appearance quality[3 pieces]

M

Structure and size[3 pieces]

M

Note: "M"-Meet the standard's requirement "F"-Fail to meet the standard's requirement
"___"-No comment

Remark:

This report is the english translation version of the report 210188492.

All the tested items are tested under the standard condition (except for indication).

Copies of the report are valid only re-stamped.

The experiment was carried out at No.1, Zhujiang Road, Panyu District, Guangzhou, Guangdong, P.R.China.

Approved By:

Yuan Liu Engineer

刘圆



Test Report

(Electronic version)

No:210188493



Test Report

(Electronic version)

No:210188493

Appearance quality[3 pieces]
(Q/STLM 001-2021 6.1)

M

Pass

Requirement

The appearance of the mask should be neat and intact, and there should be no damage or stain on the surface.

(Q/STLM 001-2021)

Structure and size[3 pieces]
(Q/STLM 001-2021 6.2)

M

Structure:

Pass

Maximum size deviation: -1%

Requirement

After wearing the mask, it should be able to cover the wearer's nose, mouth to jaw, and should conform to the design size, the maximum tolerance should not exceed $\pm 5\%$.

(Q/STLM 001-2021)



——End of Report——

Page 3 of 3